



EU REGIONAL TRUST FUND 'MADAD'

MINUTES OF THE CLARIFICATION MEETING

OUR REF: **AISPO/EU-MADAD/ME/09**

WITHIN THE FRAMEWORK OF THE PROJECT "SUPPORT TO THE EMERGENCY/CRITICAL CARE SERVICES AND MATERNAL AND CHILD HEALTH IN DUHOK GOVERNORATE"

AISPO/EU-MADAD/ME/09

Supply of biomedical equipment for Hevi Paediatric Hospital in Duhok, Autonomous Region of Kurdistan (Iraq)

Date : **18.03.2018**
Time: **9:30 am**
Venue: **Duhok Directorate General of Health**
Reference: **AISPO/EU-MADAD/ME/09**

Present:

Marco Croce (AISPO)
Amedeo Porfito (AISPO)
Sizar Ismat (AISPO)
Evren Ozkafa (DAYLIGHT Marketing Manager)

Minutes of the meeting:

AISPO representatives briefly summarised the administrative practices, stressing that:

- lots may be separately awarded to different companies;
- the maximum ceiling for the tender is Euro 299.999,99. If the total price exceeds this threshold, AISPO in collaboration with Heevi Hospital Management will decide which lots will be awarded on the basis of their prioritisation.

In addition to that, the following topic was discussed during the meeting:

- **Biomedical engineer:** for lot 21, 22, 23, 24, 25, 26, 29, 30 a biomedical engineer with a minimum of three years of experience is required for installing the biomedical equipment (as can be inferred by the professional capacity of tender selection criteria; ref: Contract Notice- paragraph 16). If a company proposes a technician for installing the aforementioned lots without the degree and/or experience required, the evaluation committee may reject the related tenderer offer/s.

The clarification meeting closed at 10:30 am.